

BOSE INSTITUTE KOLKATA

Tender No.	:	BI/Bio/PC/07/2013-14/R/5598/14
Tender date	:	05.02.2014
Type of Tender	:	Limited

Tender Title	:	ULTRASONIC BATH
Specification	:	<ol style="list-style-type: none"> 1. Bath volume 5.7- 6 Ltr 2. External and tank parts to be made of durable 304 stainless steel. 3. Temperature range Ambient to 70' C 4. Ultrasonic Power 200 (W) 5. Over-temperature protection with alarm. 6. Digital LED display using splash-proof keypads: timer (1 to 99 min, Max 4 hour operating), temperature monitor. 7. High-frequency 40 kHz sound waves 8. Minimum internal volume 295x160X160 (in mm) 9. Multiple output mode (Normal / Pulse / Sweep) with power selection (Low/Middle/ High).
Quantity	:	01 No.

Last date & time for submission	:	18.02.2014 within 1.00 p.m.
Date & time for opening of bids	:	18.02.2014 within 3.00 p.m.
Submission of Tender (Address)	:	Department of Biochemistry Bose Institute P-1/12, CIT Scheme VIIM, Kolkata 700054
Venue of bid opening	:	Same as above
For any query the interested bidders may contact (Dept./Section/Div./Unit)	:	Department of Biochemistry Bose Institute

General Terms & Conditions ::		
Warranty	:	2 years from the date of installation
Payment terms	:	Payment will be made after complete delivery of the instrument in good condition and satisfactorily installation.
Delivery schedule	:	Within 6 weeks from date of order & if any defect of the supplied item is found, it should be replaced immediately from your side.
Bid security (earnest money deposit), if applicable	:	NA

Submission of Performance Bank Guarantee (PBG), if applicable	:	NA
Any other information (if applicable)	:	
<i>Name of the instrument and submission of tender should be mentioned on the envelope positively.</i>		
Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.		

Registrar, B.I.