

BOSE INSTITUTE
KOLKATA

Tender No.	:	BI/Bio-phy/2013-14/04
Tender date	:	19.06.2013
Type of Tender	:	Open

Tender Title	:	3rd Place Balance
Specification : Weighing capacity		Up to about 200 g
Readability		1 mg. (0.001g.)
Tare range (subtractive)		about 200 g
Repeatability		± 1 mg. (0.001g.)
Linearity		± 2 mg. (0.002g.)
Pan size (diameter)		100 to 120 mm
Monolithic Weighing Technology is preferable		
In addition the instrument should have:		
Backlit LCD Display		
Glass draft shield along with a dust cover.		
Both internal and external calibration & adjustability functions		
Capability of connecting to a printer		
ISO 9001 / 14001 quality certification		
220 V AC adapter and a duster must be included with the instrument.		
Quantity	:	01 Nos.

Last date & time for submission	:	01.07.2013 not later than 1.00 p.m
Date & time for opening of bids	:	01.07.2013 to be opened at 2.30 p.m
Submission of Tender (Address)	:	Department of Biophysics, Annex Block, Centenary Building addressed to The Registrar, Bose Institute P-1/12, C.I.T. Scheme VII-M, Kolkata-700 054
Venue of bid opening	:	Department of Biophysics, Annex Block
For any query the interested bidders may contact (Dept./Section/Div./Unit)	:	Department of Biophysics Contact No. 25693336

General Terms & Conditions:	:	
Warranty	:	02 Years
Payment terms	:	Payment will be made after complete delivery of the instrument in good condition and satisfactorily installation.
Delivery schedule	:	After receipt of the work order the instrument should be delivered within fifteen days & if any defect of the supplied item is found, it should be replaced immediately from your side.
Bid security (earnest money deposit), if applicable	:	N A.
Submission of Performance Bank Guarantee (PBG), if applicable	:	N A.
Any other information (if applicable)	:	

Name of the Tender title, Tender No. & date and ***submission of tender*** should be mentioned on the envelope positively.

Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full without assigning any reasons therefor.

Registrar
19.6.2013