

BOSE INSTITUTE  
KOLKATA

Tender No.	:	R/MC/Chem/829/13
Tender date	:	12 <sup>th</sup> August 2013
Type of Tender	:	<b>Open</b>

Tender Title Specification	:	Refrigerated Circulating Water bath
	:	<ul style="list-style-type: none"> <li>• Temp. control range : -25<sup>0</sup>C to 95<sup>0</sup>C or better</li> <li>• Temp. control accuracy: + 0.1<sup>0</sup>C</li> <li>• Cooling capacity: : 200W</li> <li>• Circulation : at least 5L/min.</li> <li>• Temp. control method : Microprocessor controlled PID</li> <li>• Heater wattage : 500W</li> <li>• Parameter display/setting: Digital LED</li> <li>• Safety features : Self-diagnosis (Heater fault, sensor fault, upper/lower temperature limit alarm, low water cut out, independent over heat protector, refrigeration over load trip, high pressure cut-out, mains overload/leakage breaker</li> <li>• Aux. function : Auto start/stop with 0-99h 59 min timer, auto tuning</li> <li>• Temp. Sensor : Thermistor</li> <li>• Bath capacity : 4L</li> <li>• Operating Voltage: 220-240V, 50Hz</li> <li>• Warranty : At least 1 year</li> </ul>
Quantity	:	<b>one</b>

Last date & time for submission	:	26 <sup>th</sup> August 2013 upto 2.00 p.m.
Date & time for opening of bids	:	26 <sup>th</sup> August 2013 at 3.00PM
Submission of Tender (Address)	:	Department of Chemistry, Bose institute, Main Campus, 93/1, A.P.C. Road, Kolkata 700 009
Venue of bid opening	:	above
For any query the interested bidders may contact (Dept./Section/Div./Unit)	:	Chemistry, Main Campus, Bose institute, Ph: 2303-1143

<b>General Terms &amp; Conditions ::</b>		
Warranty	:	At least one year
Payment terms	:	Payment will be made after complete delivery of the instrument in good condition and satisfactorily installation.
Delivery schedule	:	Within 4 weeks from date of order &

Bid security (earnest money deposit), if applicable	:	if any defect of the supplied item is found, it should be replaced immediately from your side. Nil
Submission of Performance Bank Guarantee (PBG), if applicable	:	Nil
Any other information (if applicable)	:	
<i>Name of the instrument and submission of tender should be mentioned on the envelope positively.</i>		
Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.		

Registrar  
12.8.2013