

BOSE INSTITUTE
KOLKATA

| | |
|-----------------------|----------------------------|
| Tender No. | DPB/Srirupa Das/2013-14/01 |
| Tender Date | 09.09.2013 |
| Type of tender | Limited |

| | |
|---------------|---|
| Tender title | Fast blotting system with compatible electroelution facility |
| Specification | <p>(1)FAST blotting system with four gel simultaneously, 2 different protocol at same time (2) Input power:100-240 VAC,276 VA,50-60Hz, 175 W max (3) USB Port: Yes, input only for firmware updates, (4) System should have Cooling fan, (5) Should have inbuilt Power Pac supply, (6) User Interface, (7) 18 button key pad, 128 x 64 pixel monochrome display, (8) Programmable methods: Up to 25 user-defined, (9) Preprogrammed methods: Standard SD, 1.5 mm gels, High MW, Low MW, Mixed MW, (10) audible alarm: Yes, (11) User notifications should have following features: (a)Power fail during run, (b) No-load detection, (c) No cassette detection, (d) End of run.</p> <p>The Blotting system should be accompanied by a compatible electroeluter system with the following features:</p> <ul style="list-style-type: none"> • Should come with electro-eluter module, membrane caps (MW cutoff 12,000-15,000), glass tubes, frits, silicone adaptors, grommets and stoppers, buffer tank, lid with power cables, • Elution Capacity : 1 to 6 samples, collection volume: 400-600 ul |
| Quantity | One |

| | |
|---|---|
| Last date & time for submission | 20.09.2013 within 12.00 noon |
| Date & time for opening of bids | 20.09.2013 at 2.00 p.m. |
| Submission of Tender (Address) | Office of the Department of Plant Biology, (addressed to the Registrar), Bose Institute, P-1/12, C.I.T. Scheme VII M, Kolkata-700 054 |
| Venue of bid opening | Seminar Room of DPB (CB), Bose Institute |
| For any query the interested bidders may contact (Dept/Section/Div./Unit) | 033 2569 3312 |

| | |
|---|--|
| General Terms & Conditions | |
| Warranty | Minimum 2 years from the date of installation |
| Payment terms | In advance by DD or Letter of Credit |
| Delivery schedule | Within 2 months from the date of receipt the order |
| Bid security (earnest money deposit), if applicable | NA |
| Submission of performance Bank Guarantee (PBG), if applicable | NA |
| Any other information (if applicable) | NA |
| Name of the instrument and submission of tender should be mentioned on the envelop positively. | |
| <i>Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.</i> | |

Registrar
10.09.2013