

**BOSE INSTITUTE  
KOLKATA**

Tender No	<b>DPB/RM/2015-16/01</b>
Tender Date	<b>17.11.2015</b>
Tender Type	<b>Limited</b>

Tender title	<b>UV Crosslinker</b>
Specification	<ul style="list-style-type: none"> <li>• The UV Crosslinker should be especially designed for binding nucleic acids to membranes</li> <li>• The UV crosslinker should have 254nm wavelength.</li> <li>• Should include (5x8W) UV Bulbs.</li> <li>• Should highly accurate programmable microprocessor controlled photo feedback system that maintains uniform output from each of the crosslinker's five 8 watt UV bulb.</li> <li>• Exposure time should be 0-999.9 minutes.</li> <li>• Energy range should be 0-99.99 J or 0.999 J.</li> <li>• Should have UV blocking observation window.</li> <li>• Should have large visible front panel LED, with non-UV transmissible front door connected to safety interlock switches.</li> <li>• Foot print should not be more than 35x36cm.</li> <li>• Internal dimensions should be 26x33x14.5cm (WxDxH).</li> <li>• External dimensions should be 35x35x30cm (WxDxH).</li> </ul>
Quantity	1(one)
Last date & time for submission	<b>09.12.2015 upto 13.00 hrs</b>
Date & time for opening of bids	<b>09.12.2015 at 15.00 hrs</b>
Submission of Tender (Address)	<b>Department of Plant Biology, Centenary Building(CB), P-1/12, CIT Scheme VII-M, Kolkata-700054 Kind Attention : Dr. Rajeswari Mukherjee</b>
Venue of bid opening	<b>D.P.B (C.B) Room No -128</b>
For any query the interested bidders may contact (Dept/Section/ Div./Unit)	<b>Dr. Rajeswari Mukherjee 033- 25693247</b>
General Terms & Conditions	NA
Warranty	2 years from the date of installation.
Payment terms	After successful delivery and satisfactory Installation
Delivery schedule	DPB, Bose Institute, Centenary Building, P 1/12 C.I.T Scheme VII M, Kolkata-700054
Bid security (earnest money deposit), if applicable	NA
Submission of performance Bank Guarantee (PBG), if applicable	NA
Any other information (if applicable)	NA

Name of the instrument and submission of tender should be mentioned on the envelope positively.

Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons behind rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.

Sr. Professor and In-charge Registrar's Office