

BOSE INSTITUTE KOLKATA

Tender No.	:	BI/Boi-phy/2015-16/10
Tender date	:	15.02.2016
Type of Tender	:	Open

Tender Title	:	Refrigerated Micro-centrifuge.
Specification	:	<ol style="list-style-type: none"> 1. Maximum RCF should be 17,000xg. 2. Maximum speed should be above 13,200 rpm. 3. Rotor of 24x1.5ml/2ml with click seal (auto-lock) lid made of polycarbonate should be supplied with the centrifuge. The rotor should be tested and approved for biocontainment. 4. Rotor should be autoclavable. 5. Large LED display for Time, Speed and Temperature must be present. 6. Maximum noise level should not be more than 50 dBA. 7. Temperature set range must be from minus 9°C to plus 40°C. 8. Acceleration/Deceleration time must not be more than 10Sec/12 Sec. 9. Time set range 1 to 99 min with 1 min increments. 10. Toggle between RPM and RCF should be easily done. 11. Induction maintenance free rotor must be provided. 12. In future, system should be upgradable for 10x5 ml rotor with auto-lock facility. 13. One minispin centrifuge with speed 6000rpm and capacity 2ml x6 and 0.2ml x 16 should be supplied at free of cost.
Quantity	:	01 No.

Last date & time for submission	:	07.03.2016 upto 01.30 P.M
Date & time for opening of bids	:	07.03.2016 at 03.00 P.M
Submission of Tender (Address)	:	Department of Biophysics, Centenary Building, P-1/12, Cit Scheme VII-M, Kolkata 700054
Venue of bid opening	:	As above
For any query the interested bidders may contact (Dept./Section/Div./Unit)	:	033-2569 3218

General Terms & Conditions ::		
Warranty	:	02 Years.
Payment terms	:	Payment will be made after complete delivery of the instrument in good condition and satisfactorily installation
Delivery schedule	:	Within 2-3 weeks from date of order & if any defect of the supplied item is

		found, it should be replaced immediately from your side.
Bid security (earnest money deposit), if applicable	:	N.A
Submission of Performance Bank Guarantee (PBG), if applicable	:	N.A
Any other information (if applicable)	:	033-2569 3218
<i>Name of the instrument and submission of tender should be mentioned on the envelope positively.</i>		
Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.		

Sd/-
Sr. Prof. & Incharge, Registrar's Office