

बसु बिज्ञान मन्दिर

BOSE INSTITUTE

P-1/12, C.I.T. Scheme VII-M, Kolkata 700 054

BIDDING DOCUMENTS

For Tender Notice No.

BI-K/E-TEND/16/2019-20

To be addressed to:

Registrar (Officiating)
Bose Institute, Centenary Building,
P-1/12, CIT Scheme - VII -M
Kolkata - 700054 (INDIA)



Ref: BI-K/E-TEND/16/2019-20

INVITATION TO BIDS

Bose Institute, Kolkata, West Bengal, India invites *online* offers from **foreign/Indian manufacturers** or their Authorized Dealers for supply, installation and commissioning of the following equipment :

SI.	Name of the items	Qty.	Bid security (EMD)
No.			
1.	Spectral Response Measurement System	01	INR 64,000.00
2.	Glove Box	01	INR 15,000.00
			·

Details of specifications are enclosed in Annexure - I

SI.	Particulars	Date & Time
No.		
1.	Date of uploading of NIT & other documents (Online)	16.12.2019
	(Publishing date)	
2.	Documents download start date (Online)	16.12.2019 at 17:30 hrs.
3.	Clarification Start date	16.12.2019 at 17.40 hrs.
	Queries to be sent to the mail id bipurchase@jcbose.ac.in)	
4.	Clarification End date	23.12.2019 at 13:00 hrs.
5.	Prebid Conference	26.12.2019 at 15:00 hrs.
5.	Amendment (if any) (On line)	30.12.2019 at 17:00 hrs.
6.	Bid submission start date (On line)	31.12.2019 at 17:00 hrs.
7.	Documents download end date (Online)	10.01.2020 upto 17:00 hrs.
8.	Bid Submission closing (On line)	10.01.2020 upto 17:00 hrs.
9.	Last Date of submission of Earnest Money Deposit	13.01.2020 upto 17:00 hrs.
	Including the technical brochure, if any (Off line)	
10.	Bid opening date for Technical Proposals (Online)	14.01.2020 at 13:00 hrs.
11.	Date of uploading list for Technically Qualified Bidder (online)	To be notified later
12.	Date of uploading the final list of Technically Qualified Bidder	To be notified later
	(online) after disposal of appeals, if any.	
13.	Date for opening of Financial Proposal (Online)	To be notified later



Ref: BI-K/E-TEND/16/2019-20

SECTION - A

INSTRUCTION TO BIDDERS

- **A. General guidance for e-Tendering :** Instructions / Guidelines for electronic submission of the tenders have been annexed for assisting the vendors to participate in e-Tendering.
- 1. **Registration of Vendors**: Any vendor willing to take part in the process of e-Tendering will have to enrol and get registered in the Central Public Procurement (CPP) Portal, NIC, GOI through logging on to http://eprocure.gov.in/eprocure/app and the vendor is to click on the link for e-Tendering site as given on the above.
- 2. **Digital Signature certificate (DSC)**: Each vendor is required to obtain a Class-III or Class-III Digital Signature Certificate (DSC) for submission of tenders as applicable from the approved service provider of the National Informatics Centre (NIC) on payment of requisite amount. Details are available at the Web Site stated in Clause A.1. above. DSC is given as a USB e-Token.
- 3. The foreign bidders can directly submit their bid through the e-procurement by obtaining their DSC applicable from NIC, India.
- 4. The vendor can search & download N.I.T. & Tender Document(s) electronically from computer once they log on to the website mentioned in Clause A.1. using the Digital Signature Certificate. This is the only mode of collection of Tender Documents.
- 5. **Submission of Tenders**: Tenders are to be submitted through online as stated in Clause A.1. in two folders at a time for each bid, one in Technical Proposal & the other is Financial Proposal before the prescribed date & time using the Digital Signature Certificate (DSC). The documents are to be uploaded should be virus scanned copy duly Digitally Signed. The documents will get encrypted (transformed into non readable formats).
- 5. A. **Technical Proposal:** The Technical proposal should contain scanned copies of the following in two covers (folders).
 - (a) Statutory Cover containing the following documents:
 - (i) NIT (Download the NIT & upload the same using digital signature)
 - (ii) Registration Certificate for enrolment with the Central Purchase Organization / other Govt. Dept. for quoting on behalf of the foreign principals, as per latest notification of Procurement Policy Division, Department of Expenditure, Ministry of Finance.
 - (iii) Bidders' Information Form
 - (iv) Bid Form
 - (v) Techno-Commercial Bid Form
 - (vi) Manufacturer's Authorization Form
 - (vii) Bidder's Performance Statement Form
 - (viii) Service Support Details Form
 - (ix) Technical Compliance Statement Form
 - (x) Fall Clause Certificate
 - (xi) Pre Integrity Pact (for Spectral Response Measurement System)



Ref: BI-K/E-TEND/16/2019-20

(b) Non-statutory Cover containing the following documents:

For indigenous quotes the bidder has to upload requisite registration / tax certificate like TAN, PAN, Trade License, CIN, GST etc.

Note: Failure to submit any of the above mentioned documents (listed under 5(a) & (b) may render the bid liable to be summarily rejected for both statutory and non-statutory cover.

5.B. Financial Proposal:

- a. The financial bid should contain the Bill of Quantities (BOQ) corresponding to this tender, which is available on the tender webpage as an MS-Excel file.
- b. This file must be downloaded and opened with MS-Excel. The rate being quoted by the bidder must be entered in the space marked for this purpose in the file. **Quoted rate** will be encrypted under BOQ. In case quoting any rate in Tender Form, the tender is liable to be summarily rejected.
- c. All fields in the BOQ file other than those allocated for the name of the bidder and the rate(s) being quoted are non-editable.
- d. The BOQ file must be saved after this and should be uploaded using digital signature.
- e. OPTIONAL ITEMS/Additional Warranty Charge beyond Standard warranty period and training charges (if any) must be separately quoted and uploaded as PDF file.
- In the BOQ <u>quoted in foreign currency</u> (click to select currency) fill up the applicable column (i.e. 'Ex-works', 'Discount (if any)', 'Freight', 'Insurance', 'other taxes (if applicable), 'Installation & commissioning charges (if applicable)' and ignore the other column.
- In the BOQ <u>quoted in INR</u> (click to select currency) fill up the applicable column (i.e. 'Basic Rate', 'Discount (if any)', 'GST', 'Other taxes (if applicable)', 'Installation & Commissioning charges (if applicable)' and ignore the other.
- Bidders are requested not to put any percentage (%) in any column of the BOQ.
- As per Govt. Notification # 45/2017 dt.14th November, 2017, Scientific Research Institute funded by the Govt. of India, GST will be applicable @5% for the goods used for research purpose, for the bids where GST will be applicable, against DSIR Certificates to be provided by the Institute.

6. Earnest Money Deposit:

- (a). Demand Draft / Bankers Cheque / Bank Guarantee towards Earnest Money (EMD) as prescribed in the N.I.T. should in favour of "Bose Institute" payable at Kolkata.
- (b). Demand Draft / Bankers Cheque / Bank Guarantee towards Earnest Money (EMD) as prescribed in the N.I.T. against each of the instruments in favour of "Bose Institute" payable at Kolkata (original hard copy) alongwith a covering letter stating the bank details for releasing the said EMD online by the Institute, as per norms to be submitted to the Purchase Section (Import) at the Centenary Building, P-1/12, CIT Scheme VII-M, Kolkata 700054.
- (c). <u>Bidder must upload copy of valid registration certificate (i.e. MSME, NSIC) for EMD exemption.</u>



Ref: BI-K/E-TEND/16/2019-20

THE ABOVE STATED STATUTORY / NON-STATUTORY DOCUMENTS SHOULD BE ARRANGED IN THE FOLLOWING MANNER

Click the check boxes beside the necessary documents in the My Document list and then click the tab " Submit Non Statutory Documents' to send the selected documents to Non-Statutory folder.

Next Click the tab "Click to Encrypt and upload" and then click the "Technical" Folder to upload the Technical Documents.

N.B.: Failure of submission of any of the above mentioned documents as stated in Sl. No. A.5.(a). and Sl. No. A.5.(b). And 6 (a) & (b) will render the tenderer liable to be rejected for both statutory & non statutory cover.



Ref: BI-K/E-TEND/16/2019-20

- 1. Tender Evaluation Committee (TEC): Evaluation Committee constituted as per Order of the competent authority of Bose Institute, Kolkata.
- 2. Opening & evaluation of tender: If any vendor is exempted from payment of EMD, copy of relevant Government Order needs to be furnished and uploaded in the portal as PDF file.
- 3. Opening of Technical Proposal: Technical proposals will be opened either by the Sr. Prof. & Incharge, Registrar's Office / Assistant Registrar / Accounts Officer / Audit & Finance Officer, Bose Institute, Kolkata or their authorized representatives electronically from the website using their Digital Signature Certificates (DSC).
- 4. Intending tenderers may remain present if they so desire
- 5. Cover (folder) for Statutory Documents (Ref. SI. No. A.5.(a).) will be opened first and if found in order, cover (folder) for Non-Statutory Documents (Ref. SI. No. A.5.(b).) will be opened. If there is any deficiency in the Statutory Documents the bid will summarily be rejected.
- 6. Decrypted (transformed into readable formats) documents of the non-statutory cover will be downloaded & handed over to the Tender Evaluation Committee.
- 7. Pursuant to scrutiny & decision of the Tender Evaluation Committee, the summary list of eligible tenderers will be uploaded in the web portals.
- 8. During evaluation the committee may summon of the tenderers & seek clarification / information or additional documents or original hard copy of any of the documents already submitted & if these are not produced within the stipulated time frame, their proposals will be liable for rejection.

9. Financial Proposal:

- a). The vendor is to quote the rate online in the space marked for quoting rate in the BOQ. For optional items as mentioned in the specifications may be quoted separately in the additional sheet as enclosed in the Financial Cover.
- b). Only downloaded copies of the above documents are to be uploaded virus scanned & Digitally Signed by the vendor.
- c). Penalty for suppression / distortion of facts:
 - If any tenderer fails to produce the original hard copies of the documents on demand of the Tender Evaluation Committee within a specified time frame or if any deviation is detected in the hard copies from the uploaded soft copies, it may be treated as submission of false documents by the tenderer and action may be referred to the appropriate authority for prosecution as per relevant IT Act.
- (d). Rejection of Bid:
 - Director, Bose Institute reserves the right to accept or reject any Bid and to cancel the Bidding processes and reject all Bids at any time prior to the award of Contract without thereby incurring any liability to the affected Bidder or Bidders or any obligation to inform the affected Bidder or Bidders of the ground for Institute's action.
- (e). Award of Contract:

The Bidder whose Bid has been accepted finally (both technical and financial), will be informed by the Institute authority for award of contract.

The notification of award will constitute the formation of the Contract.



Ref: BI-K/E-TEND/16/2019-20

GENERAL INSTRUCTIONS

- The bids would be opened in the presence of the bidders who wish to attend the Bidopening. However, the representative should bring with them a letter of authority from the corresponding bidders, without which, they are not permitted to attend the bid opening.
- The acceptance of the quotation will rest with the competent authority of Bose Institute, Kolkata who does not bind himself to accept the lowest quotation and reserves the right to himself to reject, or partially accept any or all the quotations.
- The offer shall be valid for ninety days from the date of opening of the quotations. No revision in price will be allowed after opening the quotation.
- Unsolicited / conditional / in complete / unsigned digitally tenders shall not be considered.
- Offers that have been blindly copied from the tender specifications are not acceptable and shall be rejected.
- The Bidder / Company should not be in the list of black listed firms by any Govt. Dept. / agencies.
- The price comparison will be decided as per BOQ comparative chart on the date of opening of financial bid.
- Prices are required to be quoted in units. When quotations are given in terms of other units, relationship between two sets of units should be furnished. Quantity discounts, if any should also be indicated. The items should be quoted indicating the serial nos.
- To assess in the examination / evaluation, comparison and post qualification of the bids, purchaser may at its discretion, ask the bidder for a clarification of its bid. The request for clarification and the response shall be in writing and no change in the prices or substance of the bids shall be sought, offered or permitted. Any clarification submitted by the bidder in respect of its bid which is not in response to a request by the purchaser shall not be considered.
- Either the Indian Agent on behalf of the Foreign principal or the Foreign principal can bid directly in a tender but not both. However, the offer of the Indian Agent should also accompany the authorisation letter from their principal. To maintain sanctity of tendering system, one Indian Agent can not represent two different Foreign principals in one tender.
- Please indicate the agents in India, their address, the details of service rendered by them & the percentage of commission payable to them. In normal courses agency commission is not allowed.
- Two or more vendors cannot submit bid quoting the rates on behalf of same OEM.
- The Institute is registered with Department of Scientific & Industrial Research, Govt. of India and thus is exempted from Customs/Central Excise duty vide notification no. 10/97 dated 01.3.1997 and 16/2007 dated 01.03.2007. The present details of registration are no. 11/106/1988-TU-V dt. 31.3.2018 valid upto 31.3.2021.
- The mode of dispatch of the items must be mentioned clearly in the quotation.
- Samples, if called for, shall be submitted free of charge and or no obligation basis.
- The offered delivery period shall have to be strictly adhered to in case an order is placed.



Ref: BI-K/E-TEND/16/2019-20

SECTION B

TERMS & CONDITIONS

1. <u>Price</u>:

- Bidders are requested to give their final and best offer. <u>Techno-Commercial</u> Discussions will be held with the Lowest Bidder only, if necessary.
- Vendors, who do not accept our standard commercial terms are liable to be ignored. A
 categorical confirmation of acceptance of all our terms and conditions in toto will have to
 be observed which enable speedy processing of the offers.
- Pre-conditioned, incomplete offers, not in line with the terms and conditions of the tender documents, are liable to be rejected.

a. For Foreign Quotation:

- The price of the goods, quoted FOB/FCA port of shipment should include Export Packing, Loading Charges, Inland Freight, carriage etc. in the shipper country to be paid by the Supplier.
- The price for transportation / carriage and insurance and other costs incidental to delivery of the goods upto the port of despatch (CIF/CIP).
- Prices quoted by the bidder shall remain fixed during the entire period of contract and shall not be subject to variation on any account. A bid submitted with an adjustable price quotation will be treated as non responsive and rejected.
- Prices charged by the Supplier for Goods delivered and Services performed under the Contract shall not vary from the prices quoted by the Supplier in its bid. Conditional offers indicating changes to price quoted due to price increase / decrease by the principal firm, market fluctuations, foreign exchange fluctuations etc. shall make the bid liable to be cancelled.

b. For Indigenous Quotation:

- The price of the goods, quoted (ex-works / ex-factory / ex-showroom / ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and other taxes already paid or payable.
- The amount payable on account of GST should be mentioned clearly. If there is no
 explicit mention of taxes in your offer then quoted price will be deemed
 inclusive of such taxes. No other charges except those mentioned clearly in the
 quotation will be paid.



Ref: BI-K/E-TEND/16/2019-20

• Rates should be quoted FOR, Bose Institute, Kolkata inclusive of packing, forwarding, installation and documentation & commissioning charges etc. If ex-works prices are quoted then packing, forwarding, documentation, freight and insurance charges must be clearly mentioned separately. Vague terms like "packing, forwarding, transportation etc. extra" without mentioning the specific amount/percentage of these charges will NOT be accepted. Such offers shall be treated as incomplete and rejected. Where there is no mention of packing, forwarding, installation and commissioning, freight, documentation, insurance charges, such offers shall be summarily rejected as incomplete.

2. Period of validity of Bids:

- Bids shall remain valid at least for <u>90 days</u> after the date of bid opening prescribed by the Purchaser.
- In exceptional circumstances, the Purchaser may solicit the Bidder's consent to an extension of the period of validity.
- 3. **Bid Currencies**: Prices shall be quoted in Indian Rupees or in freely convertible foreign currency wherever possible for comparison.
- 4. Purchaser's Right to vary Quantities at the Time of Award: The Purchaser reserves the right at the time of Contract award to increase or decrease the quantity of goods and services originally specified in the Schedule of Requirements without any change in unit price or other terms and conditions. However, the Purchaser reserves the right to call the lowest firm for negotiation in case of increase in quantity.
- **Order acceptance**: The successful bidder should submit acceptance of the Purchase Order immediately but not later than 30 days in any case from the date of issue of the Purchase Order failing which it shall be presumed that the vendor is not interested and his bid security is liable to be forfeited.
- 6. **Patent Rights**: The Supplier shall indemnify the Purchaser against all third-party claims of infringement of patent, trademark or industrial design rights arising from use of the Goods or any part thereof in India.
- **7. Insurance**: The Goods supplied under the Contract shall be fully insured in Indian Rupees against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the manner
- 8. Transportation: Where the Supplier is required under the Contract to transport the Goods to a specified place of destination within India defined as Project site, transport to such place of destination in India including insurance, as shall be specified in the Contract, shall be arranged by the Supplier, and the related cost shall be included in the Contract Price.



Ref: BI-K/E-TEND/16/2019-20

9. Change Orders:

- The Purchaser may at any time, by written notice given to the Supplier make changes within the general scope of the Contract in any one or more of the following:
 - Drawings, designs, or specifications, where Goods to be furnished under the Contract are to be specifically manufactured for the Purchaser;
 - The method of shipping or packing;
 - The place of delivery; and/or
 - The Services to be provided by the Supplier.
 - The delivery schedule
 - If any such change causes an increase or decrease in the cost of, or the time required for, the Supplier's performance of any provisions under the Contract, an equitable adjustment shall be made in the Contract Price or delivery schedule, or both, and the Contract shall accordingly be amended. Any claims by the Supplier for adjustment under this clause must be asserted within fifteen (15) days from the date of the Supplier's receipt of the Purchaser's change order.

10. Penalty:

- If the Supplier fails to deliver any or all of the Goods or to perform the Services within the period(s) specified in the Contract, the Purchaser shall, without prejudice to its other remedies under the Contract, deduct from the Contract Price, as penalty, a sum equivalent to the percentage of the delivered price of the delayed Goods or unperformed Services for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the Percentage. Once the maximum is reached, the Purchaser may consider termination of the Contract.
- The applicable rate is 0.5% per week and the maximum deduction is 5% of the contract price.
- **11. Applicable Law**: The Contract shall be interpreted in accordance with the laws of the Union of India and all disputes shall be subject to place of jurisdiction at Kolkata.
- **Right to Use Defective Goods**: If after delivery, acceptance and installation and within the guarantee and warranty period, the operation or use of the goods proves to be unsatisfactory, the Purchaser shall have the right to continue to operate or use such goods until rectifications of defects, errors or omissions by repair or by partial or complete replacement is made without interfering with the Purchaser's operation.

13. Payment:

Payments shall be made promptly by the Purchaser normally after submission of the invoice or claim by the Supplier.

a. For Foreign Payment:

In case of imports, payments are usually made through Letter of Credit (LC) / wire transfer / FDD wherever applicable. The LC will be opened for 100% value but the payment shall be released partly after despatch and partly after installation & commissioning. 90% of payment will be released on



Ref: BI-K/E-TEND/16/2019-20

shipping documents and remaining 10% on successful installation and commissioning with submission of Performance Bank Guarantee of 10% of the order value. In case of Purchase Order value equivalent to INR 5 Lakh and above, 10% Performance Bank Guarantee (PBG) should be submitted before establishment of the LC/release of the payment.

b. For Indigenous Payment:

• The general conditions of payment for any indigenous items are 100% payment after successful installation & commissioning of the equipment. In case the equipment does not need any installation & commissioning the period shall be reckoned from the date of delivery & inspection. In case of Purchase Order value equivalent to INR 5 Lakh and above, 10% Performance Bank Guarantee (PBG) should be submitted before release of the payment.

14. **Delivery**:

a. For Foreign Consignment:

- Delivery of the consignment(s) should be made within a maximum of <u>8 weeks from the date of placement of purchase order unless otherwise specified so</u>. Within 24 hours of shipment, the supplier shall notify the purchaser and the insurance company by cable/telex/fax/e mail the full details of the shipment including contract number, railway receipt number/ AWB etc and date, description of goods, quantity, name of the consignee, invoice etc. The supplier shall mail the following documents to the purchaser with a copy to the insurance company:
 - 4 Copies of the Supplier invoice showing contract number, goods' description, quantity, unit price, total amount;
 - Acknowledgment of receipt of goods from the consignee(s) by the transporter;
 - Insurance Certificate if applicable;
 - Manufacturer's/Supplier's warranty certificate;
 - Inspection Certificate issued by the nominated inspection agency, if any, and the Supplier's factory inspection report;
 - Certificate of Origin; and
 - Two copies of the packing list identifying the contents of each package.
- The above documents should be received by the Purchaser before arrival of the consignment(s) (except where the Goods have been delivered directly to the Consignee with all documents) and, if not received, the Supplier will be responsible for any consequent expenses.



Ref: BI-K/E-TEND/16/2019-20

b. For Indigenous Consignment:

The delivery of the consignment(s) should be made within 4-6 weeks from the issue of the Purchase Order, unless otherwise specified.

- **15. Bank Charges : For Foreign Procurement :** The bank charges inside India to the applicant account and outside India to the beneficiary account.
- **16. Installation**: Installation should be Free of Cost and should be completed within maximum 07-10 days after delivery of the consignment.
- 17. Warranty:
 - The Supplier shall warrants that the Goods supplied under this Contract are new, unused, of the most recent or current models and that they incorporate all recent improvements in design and materials unless provided otherwise in the Contract. The Supplier shall further warrants that all Goods supplied under this Contract shall have no defect arising from design, materials or workmanship (except when the design and/or material is required by the Purchaser's Specifications) or from any act or omission of the Supplier, that may develop under normal use of the supplied Goods in the conditions prevailing in the country of final destination. The warranty should be comprehensive and on site.
 - This warranty shall remain valid at least for 12 months after the Goods or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the Contract, or for 18 months after the date of shipment whichever period concludes earlier
 - The Purchaser shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall immediately, within a reasonable period, arrange to repair or replace the defective goods or parts thereof free of cost at the ultimate destination. The Supplier shall take over the replaced parts/goods at the time of their replacement. No claim whatsoever shall lie on the Purchaser for the replaced parts/goods thereafter. In case any component has to be imported the same shall be done on DDU Kolkata basis and the purchaser shall provide the customs duty exemption certificate. If the supplier having been notified fails to remedy the defects within a reasonable period, the purchaser may proceed to take such remedial action as may be necessary, at the supplier's risk and expenses and without prejudice to any other rights, which the purchaser may have against the supplier under the contract.
 - Warranty period shall be 12 months from date of acceptance of Goods. The Supplier shall, in addition, comply with the performance and/or consumption guarantees specified under the contract. If for reasons attributable to the Supplier, these guarantees are not attained in whole or in part, the Supplier shall at its discretion make such changes, modifications, and/or additions to the Goods or any part thereof as may be necessary in order to attain the contractual guarantees specified in the Contract at its own cost and expense and to carry out further performance tests. The warranty should be comprehensive on site.
 - If a different period of warranty has been specified in the 'Technical Specifications' shall stand modified to that extent.
 - The necessary warranty certificate in this effect should be furnished along with the consignment.



Ref: BI-K/E-TEND/16/2019-20

- 18. **Training**: The Supplier is required to train the designated Purchaser's technical and end user personnel to enable them to effectively operate the total equipment.
- 19. **Supporting Equipment**: If equipment will require indigenous supporting instruments/accessories (computer, printer, online UPS etc.) at the time of the installation, the same should be quoted in separate quotation.
- 20. **Service Facility**: Supplier should mention about the possible service set up in India and how capable they are to provide after sales service.

21. Force Majeure:

- The Supplier shall not be liable for forfeiture of its performance security, liquidated damages or termination for default, if and to the extent that, its delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.
- If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
- 22. Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.



Ref: BI-K/E-TEND/16/2019-20

PRE CONTRACT INTEGRITY PACT FOR SPECTRAL RESPONSE MEASUREMENT SYSTEM

General

This pre-bid pre-contract Agre	eement (hereinafter	called the Integri	ty Pact) is made	e on	day	of of
the month of	2019, between, o	n one hand, the	Director, Bose	Institute a	cting throu	ugh
Shri	Designation	on of the officer,	Bose Institute	(hereinaft	er called	the
"BUYER", which expression	n shall mean and	include, unless	the context of	otherwise	requires,	his
successors in office and as	signs) of the First	Part and M/s		re	presented	by
Shri	_, Chief Executive	Officer (hereina	fter called the	"BIDDER/	Seller" wh	iich
expression shall mean and ir	nclude, unless the d	context otherwise	requires, his su	ccessors	and permit	ted
assigns) of the Second Part.						

WHEREAS the BUYER proposes to procure (Name of the Stores/Equipment/Item) and the BIDDER/Seller is willing to offer/has offered the stores and WHERAS the BIDDER is a private company/public company/Government undertaking/partnership/registered export agency, constituted in accordance with the relevant law in the matter and the BUYER is a Ministry/Department of the Government of India/PSU performing its functions on behalf of the President of India

NOW, THEREFORE, to avoid all forms of corruption by following a system that is fair, transparent and free from any influence/prejudiced dealings prior to, during and subsequent to the currency of the contract to be entered into with a view to:-

Enabling the BUYER to obtain the desired said stores/equipment at a competitive price in conformity with the defined specifications by avoiding the high cost and the distortionary impact of corruption on public procurement, and

Enabling BIDDERs to abstain from bribing or indulging in any corrupt practice in order to secure the contract by providing assurance to them that their competitors will also abstain from bribing and other corrupt practices and the BUYER will commit to prevent corruption, in any form, by its officials by following transparent procedures.

The parties hereto hereby agree to enter into this Integrity Pact and agree as follows:

Commitments of the BUYER

1.1 The BUYER undertakes that no official of the BUYER, connected directly or indirectly with contract, will demand, take a promise for or accept, directly or through intermediaries, any bribe, consideration, gift, reward, favour or any material or immaterial benefit or any other advantage from the BIDDER, either for themselves or for any person, organization or their party related to the contract in exchange for an advantage in the bidding process, bid evaluation, contracting or implementation process related to the contract.



Ref: BI-K/E-TEND/16/2019-20

1.1.1 The BUYER will, during pre-contract stage, treat all BIDDERs alike, and will provide to all BIDDERs the same information and will not provide any such information to any particular BIDDER which could afford an advantage to that particular BIDDER in comparison to other BIDDERS.

- 1.1.2 All the officials of the BUYER will report to the appropriate Government office any attempted or completed breaches of the above commitments as well as any substantial suspicion of such a breach.
- In case any such preceding misconduct on the part of such official(s) is reported by the BIDDER to the BUYER with full and verifiable facts and the same is prima facie found to be correct by the BUYER, necessary disciplinary proceedings, or any other action as deemed fit, including criminal proceedings may be initiated by the BUYER and such a person shall be debarred from further dealings related to the contract process. In such a case while an enquiry is being conducted by the BUYER the proceedings under the contract would not be stalled.

Commitments of BIDDERs

- 3. The BIDDER commits itself to take all measures necessary to prevent corrupt practices, unfair means and illegal activities during any stage of its bid or during any pre-contract or post-contract stage in order to secure the contract or in furtherance to secure it and in particular commit itself to the following:-
- 3.1 The BIDDER will not offer, directly or through intermediaries, any bribe, gift, consideration, reward, favour, any material or immaterial benefit or other advantage, commission, fees, brokerage or inducement to any official of the BUYER, connected directly or indirectly with the bidding process, or to any person, organization or third party related to the contract in exchange for any advantage in the bidding, evaluation, contracting and implementation of the contract.
- 3.2 The BIDDER further undertakes that it has not given, offered or promised to give, directly or indirectly any bribe, gift, consideration, reward, favour any material or immaterial benefit or other advantage, commission, fees, brokerage or inducement to any official of the BUYER or otherwise in procuring the Contract or forbearing to do or having done any act in relation to the obtaining or execution of the contract or any other contract with the Government for showing or forbearing to show favour or disfavour to any person in relation to the contract or any other contract with the Government.
- 3.3 BIDDERs shall disclose the name and address of agents and representatives and Indian BIDDERs shall disclose their foreign principals or associates.
- 3.4 BIDDERs shall disclose the payments to be made by them to agents/brokers or any other intermediary, in connection with this bid/contract.



Ref: BI-K/E-TEND/16/2019-20

- 3.5 The BIDDER further confirms and declares to the BUYER that the BIDDER is the original manufacturer/integrator/authorized government sponsored export entity of the defence stores and has not engaged any individual or firm or company whether Indian or foreign to intercede, facilitate or in any way to recommend to the BUYER or any of its functionaries, whether officially or unofficially to the award of the contract to the BIDDER, nor has any amount been paid, promised or intended to be paid to any such individual, firm or company in respect of any such intercession, facilitation or recommendation.
- 3.6 The BIDDER, either while presenting the bid or during pre-contract negotiations or before signing the contract, shall disclose any payments he has made, is committed to or intends to make to officials of the BUYER or their family members, agents, brokers or any other intermediaries in connection with the contract and the details of services agreed upon for such payments.
- 3.7 The BIDDER will not collude with other parties interested in the contract to impair the transparency, fairness and progress of the bidding process, bid evaluation, contracting and implementation of the contract.
- 3.8 The BIDDER will not accept any advantage in exchange for any corrupt practice, unfair means and illegal activities.
- 3.9 The BIDDER shall not use improperly, for purposes of competition or personal gain, or pass on to others, any information provided by the BUYER as part of the business relationship, regarding plans, technical proposals and business details, including information contained in any electronic data carrier. The BIDDER also undertakes to exercise due and adequate care lest any such information is divulged.
- 3.10 The BIDDER commits to refrain from giving any complaint directly or through any other manner without supporting it with full and verifiable facts.
- 3.11 The BIDDER shall not instigate or cause to instigate any third person to commit any of the actions mentioned above.
- 3.12 If the BIDDER or any employee of the BIDDER or any person acting on behalf of the BIDDER, either directly or indirectly, is a relative of any of the officers of the BUYER, or alternatively, if any relative of an officer of the BUYER has financial interest/stake in the BIDDER's firm, the same shall be disclosed by the BIDDER at the time of filing of tender.
 - The term 'relative' for this purpose would be as defined in Section 6 of the Companies Act 1956.
- 3.13 The BIDDER shall not lend to or borrow any money from or enter into any monetary dealings or transactions, directly or indirectly, with any employee of the BUYER.



Ref: BI-K/E-TEND/16/2019-20

4. Previous Transgression

- 4.1 The BIDDER declares that no previous transgression occurred in the last three years immediately before signing of this Integrity Pact, with any other company in any country in respect of any corrupt practices envisaged hereunder or with any Public Section Enterprise in India or any Government Department in India that could justify BIDDER's exclusion from the tender process.
- 4.2 The BIDDER agrees that if it makes incorrect statement on this subject, BIDDER can be disqualified from the tender process or the contract, if already awarded, can be terminated for such reason.

5. Earnest Money (Security Deposit)

- 5.1 While submitting commercial bid, the BIDDER shall deposit an amount of □20,10,000.00 as Earnest Money/Security Deposit, with the BUYER through any of the following instruments:
 - (i) Bank Draft or a Pay Order in favour of Bose Institute, Kolkata.
 - (ii) A confirmed guarantee by an Indian Nationalised Bank, promising payment of the guaranteed sum to the BUYER on demand within three working days without any demur whatsoever and without seeking any reasons whatsoever. The demand for payment by the BUYER shall be treated as conclusive proof of payment.
 - (iii) Any other mode or through any other instrument such as NEFT/RTGS.
- 5.2 The Earnest Money/Security Deposit shall be valid upto a period of two years or the complete conclusion of the contractual obligations to the complete satisfaction of both the BIDDER and the BUYER, including warranty period, whichever is later.
- 5.3 In case of the successful BIDDER a clause would also be incorporated in the Article pertaining to Performance Bond in the Purchase Contract that the provisions of Sanctions for Violation shall be applicable for forfeiture of Performance Bond in case of a decision by the BUYER to forfeit the same without assigning any reason for imposing sanction for violation of this Pact.
- 5.4 No interest shall be payable by the BUYER to the BIDDER on Earnest Money /Security Deposit for the period of its currency.

6. Sanctions for Violations

6.1 Any breach of the aforesaid provisions by the BIDDER or any one employed by it or acting on its behalf (whether with or without the knowledge of the BIDDER) shall entitle the BUYER to take all or any one of the following actions, wherever required:-



Ref: BI-K/E-TEND/16/2019-20

(i) To immediately call off the pre contract negotiations without assigning any reason or giving any compensation to the BIDDER. However, the proceedings with the other BIDDER(s) would continue.

- (ii) The Earnest Money Deposit (in pre-contract stage) and/or Security Deposit/Performance Bond (after the contract is signed) shall stand forfeited either fully or partially, as decided by the BUYER and the BUYER shall not be required to assign any reason therefore.
- (iii) To immediately cancel the contract, if already signed, without giving any compensation to the BIDDER.
- (iv) To recover all sums already paid by the BUYER, and in case of an Indian BIDDER with interest thereon at 2% higher than the prevailing Prime Lending Rate of State Bank of India, while in case of a BIDDER from a country other than India with interest thereon at 2% higher than the LIBOR. If any outstanding payment is due to the BIDDER from the BUYER in connection with any other contract for any other stores, such outstanding payment could also be utilized to recover the aforesaid sum and interest.
- (v) To encash the advance bank guarantee and performance bond/warranty bond, if furnished by the BIDDER, in order to recover the payments, already made by the BUYER, along with interest.
- (vi) To cancel all or any other Contracts with the BIDDER. The BIDDER shall be liable to pay compensation for any loss or damage to the BUYER resulting from such
- (vii) cancellation/rescission and the BUYER shall be entitled to deduct the amount so payable from the money(s) due to the BIDDER.
- (viii) To debar the BIDDER from participating in future bidding processes of the Government of India for a minimum period of five years, which may be further extended at the discretion of the BUYER.
- (ix) To recover all sums paid in violation of this Pact by BIDDER(s) to any middleman or agent or broker with a view to securing the contract.
- (x) In cases where irrevocable Letters of Credit have been received in respect of any contract signed by the BUYER with the BIDDER, the same shall not be opened.
- (xi) Forfeiture of Performance Bond in case of a decision by the BUYER to forfeit the same without assigning any reason for imposing sanction for violation of this Pact.
- 6.2 The BUYER will be entitled to take all or any of the actions mentioned at Para 6.1(i) to (x) of this Pact also on the Commission by the BIDDER or any one employed by it or acting on its behalf (whether with or without the knowledge of the BIDDER), of an offence as defined in Chapter IX of the Indian Penal Code, 1860 or Prevention of Corruption Act, 1988, or any other statute enacted for prevention of corruption.



Ref: BI-K/E-TEND/16/2019-20

6.3 The decision of the BUYER to the effect that a breach of the provisions of this Pact has been committed by the BIDDER shall be final and conclusive on the BIDDER. However, the BIDDER can approach the independent Monitor(s) appointed for the purposes of this Pact.

7. Fall Clause

7.1 The BIDDER undertakes that it has not supplied/is not supplying similar product/systems or subsystems at a price lower than that offered in the present bid in respect of any other Ministry/Department of the Government of India or PSU and if it is found at any stage that similar product/systems or sub systems was supplied by the BIDDER to any other Ministry/Department of the Government of India or a PSU at a lower price, then that very price, with due allowance for elapsed time, will be applicable to the present case and the difference in the cost would be refunded by the BIDDER to the BUYER, if the contract has already been concluded.

8. Independent Monitors

- 8.1 The BUYER has appointed independent Monitors (hereinafter referred to as Monitors) for this Pact in consultation with the Central Vigilance Commission (Independent External Monitors appointed by CVC for the tender -Dr. Sanjay Kumar Panda, IAS (Retd.) & Shri Amol Prabhakar Joshi, CES (Retd.)
- 8.2 The task of the Monitors shall be to review independently and objectively, whether and to what extent the parties comply with the obligations under this Pact.
- 8.3 The Monitors shall not be subject to instructions by the representatives of the parties and perform their functions neutrally and independently.
- 8.4 Both the parties accept that the Monitors have the right to access all the documents relating to the project/procurement, including minutes of meetings.
- 8.5 As soon as the Monitor notices, or has reason to believe, a violation of this Pact, he will so inform the Authority designated by the BUYER.
- 8.6 The BIDDER(s) accepts that the Monitor has the right to access without restriction to all Project documentation of the BUYER including that provided by the BIDDER. The BIDDER will also grant the Monitor, upon his request and demonstration of a valid interest, unrestricted and unconditional access to his project documentation. The same is applicable to Subcontractors. The Monitor shall be under contractual obligation to treat the information and documents of the BIDDER/Subcontractor(s) with confidentiality.
- 8.7 The BUYER will provide to the Monitor sufficient information about all meetings among the parties related to the Project provided such meetings could have an impact on the contractual relations between the parties. The parties will offer to the Monitor the option to participate in such meetings.



D. C. DI V.E. TEND (4/ 10040-00)

Ref: BI-K/E-TEND/16/2019-20

The Monitor will submit a written report to the designated Authority or BUYER/Secretary in the Department/within 8 to 10 weeks from the date of reference or intimation to him by the BUYER / BIDDER and should the occasion arise, submit proposals for correcting problematic situations.

9. Facilitation of Investigation

In case of any allegation of violation of any provisions of this Pact or payment of commission, the BUYER or its agencies shall be entitled to examine all the documents including the Books of Accounts of the BIDDER and the BIDDER shall provide necessary information and documents in English and shall extend all possible help for the purpose of such examination.

10. Law and Place of Jurisdiction

This Pact is subject to Indian Law. All disputes arising our of this Contract will be subjected to Kolkata jurisdiction.

11. Other Legal Actions

The actions stipulated in this Integrity Pact are without prejudice to any other legal action that may follow in accordance with the provisions of the extant law in force relating to any civil or criminal proceedings.

12. Validity

- 12.1 The validity of this Integrity Pact shall be from date of its signing and extend up to 5 years or the complete execution of the contract to the satisfaction of both the BUYER and the BIDDER/Seller, including warranty period, whichever is later. In case BIDDER is unsuccessful, this Integrity Pact shall expire after six months from the date of the signing of the contract.
- 12.2 Should one or several provisions of this Pact turn out to be invalid; the remainder of this Pact shall remain valid. In this case, the parties will strive to come to an agreement to their original intentions.

	intentions.	
13.	The parties hereby sign this Integrity Pact at	on
	BUYER	BIDDER
	Name of the Officer:	CHIEF EXECUTIVE OFFICER
	Designation	
	BOSE INSTITUTE	
	Witness	<u>Witness</u>
	1	1
	2	2

^{*}Provisions of these clauses would need to be amended / deleted in line with the policy of the BUYER in regard to involvement of Indian agents of foreign suppliers.

^{*} The bidders may note that they must upload the Integrity Pact signed by both the bidder as well as the buyer (Institute) and upload the same alongwith the bidding document.



Ref: BI-K/E-TEND/16/2019-20

Bidder Information Form

[The Bidder shall fill in this Form in accordance with the instructions indicated below. No alterations to its format shall be permitted and no substitutions shall be accepted. This should be done of the letter head of the firm]

Date: [insert date of Bid Submission] Tender No.: [insert number for Invitation for bids]

- 1. Bidder's Legal Name [insert Bidder's legal name]
- 2. In case of JV, legal name of each party: [insert legal name of each party in JV]
- 3. Bidder's actual or intended Country of Registration: [insert actual or intended Country of Registration]
- 4. Bidder's Year of Registration: [insert Bidder's year of registration]
- 5. Bidder's Legal Address in Country of Registration: [insert Bidder's legal address in country of registration]
- 6. Bidder's Authorized Representative Information

Name: [insert Authorized Representative's name]
Address: [insert Authorized Representative's Address]

Telephone/Fax numbers: [insert Authorized Representative's telephone/fax numbers]

Email Address: [insert Authorized Representative's email address]

7. Attached are copies of original documents of: [check the box(es) of the attached original documents]

Signature of the Bidder Name Business Address



Ref: BI-K/E-TEND/16/2019-20

BID FORM

(to be submitted on the official letterhead of the bidder)

The Director, Bose Institute, P-1/12, CIT Scheme-VII/M, Kolkata - 700054 (W.B.) India

Sir, Having examined the bidding documents the receipt of which is hereby duly acknowledged, I/We agree to furnish required supplies/services in conformity with the Techno-Commercial Bid or such portions thereof at the prices given in the Price Bid on receipt of order for the same.
I/We agree to hold this offer open untiland to supply, install and commission the equipment and complete the whole of the work and hand over to the purchaser within the period of weeks, from the date of receipt of intimation from you regarding acceptance of this tender/receipt of supply order.
I/We agree to submit the bank guarantee as specified in the form prescribed by your goodself for the due performance of the contract, if our bid is accepted.
I/We understand that you are not bound to accept the lowest or any bid you may receive.
() Signature of Bidder With Office Stamp Name & Address
Telephone No.:
FAX No.:
E-mail address:
Name & Designation of the Contact Person:



Ref: BI-K/E-TEND/16/2019-20

TECHNO-COMMERCIAL BID FORM

1.	Tender reference No	:	
2.	Due date of bid submission	:	
3.	Main Item Model No.	:	
4.	Life Expectancy of the Equipment (in years)	:	
5.	Warranty Period (in years)	:	
6.	Delivery Period of the Equipment	:	
7.	Details of Bank Guarantee enclosed as Bid Security:	:	
	Name of the Bank	:	
	Branch	:	
	Address with Phone, FAX & E-mail Nos.	:	
	Amount Rs.	:	
	Valid Upto	:	
8.	Country of the Origin of the Equipment	:	
9.	Bid currency	:	

Contd.....



Ref: BI-K/E-TEND/16/2019-20

10.	Schedule of Requirements :	:	
I.	Whether agreeing to all the terms and conditions including payment terms as		
	mentioned in the bidding documents		
II.	Port of Shipment		
III.	Approximate Shipment Weight (chargeable		
	weight) in Kg. of the packed consignment		
IV.	Approximate Dimensions/ Volume of the		
	packed consignment		
Note:			
1.	Adhering to the format given above is a		
	pre-requisite for considering your bid.		
_			
2.	All columns must be filled up.		
0			
3.	Separate list should be attached where		
	required in the same format giving details		
_	of each item.		
4.	Please indicate applicability.		
1./		<u> </u>	
	certify that I/We have completely read ar	na u	nderstood and agree to all the terms &
condit	ions aiven in ITB.		

Signature of Bidder With Office Stamp Name & Address



Ref: BI-K/E-TEND/16/2019-20 **MANUFACTURERS' AUTHORIZATION FORM** Dated _____ No. _____ The Director, Bose Institute, P-1/12, CIT Scheme-VII/M, Kolkata - 700054 (W.B.) India Dear Sir: No company or firm or individual other than M/s. ______ is authorized to bid, and conclude the contract in regard to this business. I/We hereby extend our full guarantee and warranty as per Clause 15 of the General Conditions of Contract and Clause of the Special Conditions of Contract for the goods and services offered by the above firm. Yours faithfully, (Name)

Note: This letter of authority should be on the <u>letterhead of the manufacturer</u> and should be signed by a person competent and having the power of attorney to bind the manufacturer. It should be included by the Bidder in its techno-commercial unpriced bid.

(Name of manufacturers)



Ref: BI-K/E-TEND/16/2019-20

Name of the Firm.....

BIDDER'S PERFORMANCE STATEMENT FORM

(For a Period of Last 3 Years)

Order placed by (full address with Tel. No., Fax No. & e-mail address of purchaser)	Date	Description and quantity of ordered equipment	Date of completion of delivery as per Contract/ Actual	Remarks indicating reasons for late delivery, if any	Has the equipment been installed satisfactory? (Attach a certificate from the purchaser/ Consignee)

Place:	Signature :
Date :	Office stamp



Ref : BI-K/E-TEND/16/2019-20

SERVICE SUPPORT DETAILS FORM

SI. No.	Nature of training imparted	List of similar type equipments serviced in the past 3 years	Address, Telephone Nos., Fax and e mail address of the firm located in Kolkata, if any	Value of minimum stock of consumable spares held at all times.

	Signature and Seal of the manufacturer/Bidder
Place :	
Date :	



Ref: BI-K/E-TEND/16/2019-20

TECHNCAL COMPLIANCE STATEMENT FORM

An item-by-item commentary on the Purchaser's Technical Specifications demonstrating substantial responsiveness of the goods and services to those specifications or a statement of deviations and exceptions to the provisions of the Technical Specifications.

ITEM NA	TEM NAME					
SI.No.	Tender Specifications	Bidder's Specifications	Deviation/ Remarks, if any Including justification			

(Technical literature/brochures/manuals should be attached along with this format)

Please note:

- 1. Compliance/Deviation statement comparing the specifications of the quoted model to the required specifications. This statement should also give the page number(s) of the technical literature where the relevant specification is mentioned.
- 2. Bids must have supporting documents (technical literature or copies of relevant pages from the service manual or factory test data) for all the points noted above, failure regarding which may result in rejection of bid.
- 3. In case the bidder furnishes wrong or false information wilfully in the technical compliance chart, such action shall be viewed strictly and the bidders may be blacklisted.
- 4. Where there is no deviation, the statement should be returned duly signed with an endorsement indicating "No Deviations".

		Signature and Seal of the manufacturer/Bidder
Place	:	9
Date	:	



Ref: BI-K/E-TEND/16/2019-20	
Quotation No	Date :
PROFORMA OF FALL CLAUS	SE CERTIFICATE
If on any subsequent date after submission of quotat (the term manufacturer will also include his authoriz such stores or sells such stores to any party at a p against supply order placed by Bose Institute, Kolka distributor / agent) as aforesaid in case the quotation effected by them) will forth-with notify such reduction payable for the stores to be supplied against the Supp price coming into force shall reduced correspondingly	ted distributor / agent) reduces the sale price of rice lower than the price charged / chargeable ata, the manufacturer (including his authorized is submitted by them and supply / service is also in in sale price to Bose Institute, Kolkata and price oly Order after the date of such reduction in sale
	(Signature & Date of Bidders with Rubber Stamp)
s certificate should be uploaded on the letter head of a	the company / firm / agency



Ref: BI-K/E-TEND/16/2019-20

BID SECURITY FORM

(To be executed by a scheduled Bank in India on non-judicial stamp paper)

To The Director, Bose Institute, P-1/12, CIT Scheme-VII/M, Kolkata - 700054 (W.B.) India

1.

2.

Kolkata - 700054 (W.B.) India
Dear Sir,
Whereas
(Name of country), having our registered office at
Purchaser") for the sum of
THE CONDITIONS of this obligation are:
If the Bidder withdraws its Bid during the period of bid validity specified by the Bidder on the Bid Form; or If the Bidder, having been notified of the acceptance of its bid by the Purchaser during the period of bid validity:
contd



Ref: BI-K/E-TEND/16/2019-20 Fails or refuses to execute the Contract Form if required; or 3. Fails or refuses to furnish the performance security, in accordance with the Instruction to 4. Bidders. 5. This guarantee is valid up to(six months from date). We undertake to pay the Purchaser up to the above amount upon receipt of its first written demand, without the Purchaser having to substantiate its demand, provided that in its demand the Purchaser will note that the amount claimed by it is due to it, owing to the occurrence of one or both of the two conditions, specifying the occurred condition or conditions. This guarantee will remain in force up to and including forty-five (45) days after the period of the bid validity, and any demand in respect thereof should reach the Bank not later than the above date. Date: Signature..... Place: Seal of Bankers (Complete Address/Contact Details with Tel./Fax/email etc) Witness: 1. 2.



Ref: BI-K/E-TEND/16/2019-20

per the contract during warranty period.

1.

2.

PERFORMANCE SECURITY FORM

(To be executed by a scheduled Bank in India on non-judicial stamp paper)

To The Director, Bose Institute, P-1/12, CIT Scheme-VII/M, Kolkata - 700054 (W.B.) India Dear Sir, Sub: Your Contract No. ------for You have entered into a contract with reference no as given above with __ (herein after referred to as the contractor) for the supply, installation, and commissioning of (herein after referred to as stores) for the price and on the terms and conditions contained in the said contract. In accordance with the terms of said contract, the contractor has undertaken to produce a bank quarantee for Rs. _____ (Rupees ____ ____ only) being 10 % of the total value of the said stores supplied to you, for the due fulfilment of its obligations to you for due performance as

- 3. In consideration thereof, we hereby expressly, irrevocably and unconditionally undertake and guarantee as principal obligator on behalf of the contractor that in the event you submit a written demand to us that the contractor has not performed according to the contractual obligations included in the said contract, we will pay you on written demand, without demur and without reference to the contractor any sum up to a maximum amount of Rs ______ (Rupees ______ only). Your demand shall be conclusive evidence to us that such payment is due under the terms of the said contract. Payment by us to you will be made within thirty (30) days from receipt of your request making reference to this guarantee and on demand.
- 4. This guarantee shall not be revoked without your express consent and shall not be affected by your granting any indulgence to the contractor, which shall include but not be limited to postponement from time to time of the exercise of any powers vested in you or any right which you may have against the contractor and to exercise the same in any manner at any time and either to forbear or to enforce any covenant contained or implied in the said contract or any other course or remedy or security available to you, and our Bank shall not be released from its obligations under this guarantee by your exercising any of your rights with reference to matters aforesaid or any of them or by reason of any other act or forbearance or other acts of omission or commission on your part or any other indulgence shown by you or by any other matter or thing whatsoever which under law would, but for this provision, have the effect of relieving our Bank from its obligation under this guarantee.



Ref: BI-K/E-TEND/16/2019-20

5.	Notwithstanding anything herein contained, our liability under this guarantee is restricted to Rs (Rupees only) and the guarantee shall remain in force up to and including the day of being reported to us by you and returned to us duly discharged.		
6.	Unless a demand or claim under this guarantee is made on us in writing on or before the aforesaid expiry date as provided above or unless this guarantee is extended by us all your rights under this guarantee shall be proscribed and we shall be discharged from the liabilities hereunder.		
7.	This guarantee shall not be affected by any change in the constitution of our Bank or of the contracto or for any other reason whatsoever.		
Date:	Signature		
Place:	Seal of Bankers		
	(Complete Address/Contact Details with Tel./Fax/email etc)		
	Witness:		
	1		
	2		



Ref: BI-K/E-TEND/16/2019-20

	Technical Specification of Spectral Response Measurement System		
SL. NO.	PARAMETERS	SPECIFICATIONS / REQUIREMENTS	
	Scope of measurement	Spectral Response of Infrared Photodetector.	
		System to be capable to perform DC	
		measurements 250- 1700 nm and AC	
		measurements 250-2400 nm or above.	
	Scope of Supply	Supply, installation, commissioning and training	
	Pre-qualification criteria	The supplier must have supplied similar	
		equipment to any national or international	
		laboratories. Documentary evidence must be	
		provided with contact details.	
	Commissioning	Complete set up including the interconnection of	
		all system components with necessary	
		connectors, cables, probes, etc should be	
		provided by the supplier	
		The supplier should provide a demonstration with	
		a sample on completion of successful	
		commissioning.	
	i) INFRARED PHOTODETECTOR MEASUREMENT SYSTEM 1. Monochromator		
1.1	Monochromator configuration	Czerny Turner configuration with filter wheels	
	3	3	
1.2	Monochromator focal length	≥ 300mm	
1.3	Slit width & height	Alit size at least 20mm in Y direction and variable	
		in X -direction with atleast.01-5mm size.	
		Slits should be adjustable and square/ rectangular	
		spots should be emitted from exit port of	
		monochromator varying from ~1mm-4mm.	
1.4	Bandwidth	Adjustable, measurements will be typically done	
		at 5 nm.	

1.5	Gratings	3 gratings:
1.5	Gratings	2400I/mm blazed @ UV
		1200 I/mm blazed @ Visible
		600 I/mm blazed @ IR
1.6	Grating Size	≥68mm
1.7	Stray light rejection at 10x FWHM	5·10 ⁻⁵
1.8	Wavelength coverage	250-2400 nm or above
1.11	f-number	f/4.1 or faster
1.12	Wavelength accuracy	$\leq \pm 0.05\%$ ($\leq \pm 0.2$ nm) or better
	j ,	with 1200g/mm grating at
		500 nm
1.14	Long pass &Bandpass filters	Suitable combination of long pass and bandpass
		filters to reject and allow light for high SNR.
2.	Detectors (All detectors should have	$NEP < 9 \times 10^{-11} \text{ W/ H}^{1/2}$
2.1	Silicon photodiode,(200-1100nm). NMI	Material: Silicon
	traceable calibration certificate to be	Active area≥10x10 mm
	provided.	Spectral response range : 200-1100nm
		Dark current <1pA
2.2	Ge Photodiode (800-1800nm). NMI	Active area≥ 5 mm diameter
	traceable calibration certificate to be	Material: Germanium
	provided.	Spectral response range: 800-1800nm or better
		Dark current (typ.): <1pA
2.4	Lead sulphide (1-3 um) with suitable	Material: Lead Sulphide
	Peltier cooling mechanism to minimize	Spectral response range : 1000-3000nm
	noise. Controller and accessories for TEC	Active Area ≥ 3X3 mm
	to be provided. Suitable High Voltage	Detector Temperature ≤ -10° C
	power supply for detector to be	HV power supply to have ≤ 100 ppm p-p noise and
	provided	temperature stability
3.	Light Source	
3.1	Dual xenon/ quartz halogen light Source	Lamp type: Dual xenon/ quartz halogen light source
	suitable to operate between 250-3000nm.	with facility for high and low frequency optical
		chopping with suitable light collection optics.
		Automated lamp selection using mirror assembly on
		setting wavelengths
		Xe ≥ 75W
		QTH ≥ 100W
4.	Electronics	
4.2	Optical Chopper & Control Module	Chopping Frequency: Atleast 25 Hz to 1kHz
		Phase Jitter ≤ :1° pk-pk over entire frequency range.
		Reference Output : Suitable sensor to provide
		frequency feedback to lock-in amplifier.
		Resolution1Hz or better
		Temperature Dependence of Frequency Stability:
		Less than 0.01% change per K
		Line Voltage Dependence of Frequency Stability:

Iline voltage: 4.3 AC Current Pre-amplifier Gain Ranges: Atleast 10 ⁸ -10 ³ V/A Gain Accuracy≥±1% Gain Stability≥200ppm/*C Frequency RangeAtleast10Hz - 2kHz Gain Accuracy≥±0.15% Gain Stability≤200ppm/*C ADC Resolution ≥16 bit ADC Speed≥52 kHz Programmable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current Power Supply (two numbers, one for each light source) Operating voltage: 230V, 5A, 50Hz Output voltage & current: as per system requirement Resolution≤0.1A local Output power: At least 50W-250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥±(0.04% +0.05A) Drift≤0.05% of rated output Mode types: Constant current, constant vol Front panel control features: Power on/loff select pre-set current, define current/ power Display: LCD display of set current, actual cual power Compliance to standards: EN 61010-1:2010 61326-1 Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. Section	rom a chango in	Less than 0.01% change results from a char		
4.3 AC Current Pre-amplifier Gain Ranges: Atleast 10 ⁸ -10 ³ V/A Gain Accuracy ≥+1% Gain Stability ≥200ppm/°C Frequency RangeAtleast10Hz - 2kHz Gain Accuracy ≥±0.15% Gain Stability ≤ 200ppm/°C ADC Resolution ≥16 bit ADC Speed ≥ 52 kHz 5. Power Supply 5.1 Programmable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current Power Supply (two numbers, one for each light source) Output voltage & current: as per system requirement Resolution≤ 0.1A local Output power: At least 50W-250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant voil Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual or lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 7. Optic Relay optics Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. Spectral acquisition software and measure guidance, calculation and reporting applic software should be provided. Should be	rom a change in	S .		
Gain Accuracy ≥+1% Gain Stability ≥200ppm/°C	<u> </u>	9	AC Current Dro amplifier	12
Gain Stability ≥200ppm/°C 4.4 DSP Lock-In Amplifier Frequency RangeAtleast10Hz - 2kHz Gain Accuracy ≥±0.15% Gain Stability ≥200ppm/°C ADC Resolution ≥16 bit ADC Speed ≥ 52 kHz 5. Power Supply 5.1 Programmable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current Power Supply (two numbers, one for each light source) Operating voltage: 230V, 5A, 50Hz Output voltage & current: as per system requirement Resolution≤ 0.1A local Output power: At least 50W-250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant volent programmable provided specific pre-set current, define current/ power Display: LCD display of set current, actual culamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. Spectral acquisition software and measur guidance, calculation and reporting applic software should be provided. Should be	1		Ac current Fre-ampliner	4.3
SPP Lock-In Amplifier Frequency RangeAtleast10Hz - 2kHz Gain Accuracy ≥±0.15% Gain Stability ≤ 200ppm/°C ADC Resolution ≥16 bit ADC Speed ≥ 52 kHz		3		
Gain Accuracy ≥±0.15% Gain Stability ≤ 200ppm/*C ADC Resolution ≥16 bit ADC Speed ≥ 52 kHz 5. Power Supply 5.1 Programmable Constant Current Power Supply (two numbers, one for each light source) Frogrammable Constant Current Power Supply (two numbers, one for each light source) Operating voltage: 230V, 5A, 50Hz Output voltage & current: as per system requirement Resolution≤ 0.1A local Output power: At least 50W-250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant volent features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual of lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisitient for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting applic software should be provided. Should be software should be provided. Should be software should be provided. Should be	LI I =		DCD Look in Amplifier	1.1
Gain Stability ≤ 200ppm/°C ADC Resolution ≥16 bit ADC Speed ≥ 52 kHz 5. Power Supply 5.1 Programmable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current Power Supply (two numbers, one for each light source) Output voltage & current: as per system requirement Resolution≤ 0.1A local Output power: At least 50W-250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant volf-ront panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual or lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisitification and obtain high beam unifor and beam power. Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. Spectral acquisition software and measur guidance, calculation and reporting applic software should be provided. Should be	KHZ		DSP Lock-in Amplilier	4.4
ADC Resolution ≥16 bit ADC Speed ≥ 52 kHz	_	3		
5. Power Supply 5.1 Programmable Constant Current Power Supply (two numbers, one for each light source) Frogrammable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current: as per system requirement Resolution≤ 0.1A local Output power: At least 50W- 250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, define current/ power on/off select pre-set current, define current/ power Display: LCD display of set current, accurate to gisplay: LCD display of set current, accura				
5. Power Supply 5.1 Programmable Constant Current Power Supply (two numbers, one for each light source) Programmable Constant Current Power Supply (two numbers, one for each light source) Output voltage & current: as per system requirement Resolution≤ 0.1A local Output power: At least 50W-250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant voited Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual cultamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisition. 7. Optic Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. Spectral acquisition software and measure guidance, calculation and reporting applic software should be provided. Should be	_			
Software Centrel Supply (two numbers, one for each light source) Operating voltage: 230V, 5A, 50Hz		ADC Speed ≥ 52 kHz		
Supply (two numbers, one for each light source) Supply (two numbers, one for each light source)			Power Supply	5.
source) requirement	Ηz	Operating voltage: 230V, 5A, 50Hz	Programmable Constant Current Power	5.1
Resolution≤ 0.1A local Output power: At least 50W- 250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant vo Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual of lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisitis 7. Optic Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measur guidance, calculation and reporting applic software should be provided. Should be	system	Output voltage & current: as per system	Supply (two numbers, one for each light	
Output power: At least 50W- 250W Operating conditions: Ambient temperature +40 °C Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant voil Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual of lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisitistically. 7. Optic Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be		requirement	source)	
Operating conditions: Ambient temperature +40 °C Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant vol Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual of lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any f aberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measur guidance, calculation and reporting applic software should be provided. Should be		Resolution≤ 0.1A local		
+40 °C Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant void Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual cull lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting applic software should be provided. Should be	OW O	Output power: At least 50W- 250W		
Accuracy: ≥ ±(0.04% +0.05A) Drift ≤ 0.05% of rated output Mode types: Constant current, constant void Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual cull lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiting. 7. Optic Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measuring guidance, calculation and reporting application software should be provided. Should be	emperature: 0°C -	Operating conditions: Ambient temperatur		
Drift ≤ 0.05% of rated output Mode types: Constant current, constant vol Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual of lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting applic software should be provided. Should be		+40 °C		
Mode types: Constant current, constant volume Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual culamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be		Accuracy: ≥ ±(0.04% +0.05A)		
Front panel control features: Power on/off select pre-set current, define current/ power Display: LCD display of set current, actual or lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic 7.1 Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting applic software should be provided. Should be		Drift ≤ 0.05% of rated output		
select pre-set current, define current/ power Display: LCD display of set current, actual or lamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be	onstant voltage	Mode types: Constant current, constant vo		
Display: LCD display of set current, actual collamp voltage, burn time and power Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measuring guidance, calculation and reporting application software should be provided. Should be	ver on/off lamp,	Front panel control features: Power on/off		
lamp voltage, burn time and power	rent/ power	select pre-set current, define current/ power		
Compliance to standards: EN 61010-1:2010 61326-1 6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be	nt, actual current,	Display: LCD display of set current, actual co		
6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting applications of tware should be provided. Should be	ver	lamp voltage, burn time and power		
6. Source Meter 6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Relay optics Relay optics Relay optics Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be)10-1:2010, EN	Compliance to standards: EN 61010-1:2010		
6.1 Keithley 2450 Source Meter Unit with software integration for control and data acquisiti 7. Optic Relay optics Relay optics Relay optics Relay optics Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be		61326-1		
7. Optic Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be			Source Meter	6.
Reflective relay optics are needed to be for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be	ta acquisition.	ware integration for control and data acquisiti	Keithley 2450 Source Meter Unit with soft	6.1
7.1 Relay optics for the entire wavelength to avoid any faberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be			Optic	7.
aberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be	ded to be usable	Reflective relay optics are needed to be		
aberration and obtain high beam unifor and beam power. 8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be	,		Relay ontics	7 1
8. Software Spectral acquisition software and measure guidance, calculation and reporting application software should be provided. Should be	am uniformity		Relay optics	7.1
Spectral acquisition software and measure guidance, calculation and reporting application software Software Control. 8.1 Software Control.	_	and beam power.		
guidance, calculation and reporting applic software should be provided. Should be			Software	8.
Compatible with Windows7and 8. Include Excel export and simple integration of use DLLs	rting application Should be d 8. Includes direct	guidance, calculation and reporting applic software should be provided. Should be Compatible with Windows7and 8. Include Excel export and simple integration of use	Software Control	8.1

9.	Wa	rranty	: One	year
----	----	--------	-------	------

Note: Do not prepare the compliance certificate by just copy and pasting the above specifications. Bidder must specify their own specifications against all the points including supporting software section in their own words. Original OEM Brochure should be highlighted to support each specification. Same original brochure should be downloadable from OEM website. If information is not given in the brochure, a separate authentication certificate from OEM should be attached. Copying and pasting specifications and just writing YES/ MEET the specification, may lead to rejection of the Bid.

Vendor must provide complete compliance statement against each technical point in below given format. Failing to this may leads to rejection of the Bid.

SL. NO	TENDER REQUESTED SPECIFICATION	VENDOR'S OFFERED SPECIFICATION	ANY DEVIATION FROM REQUIRED SPECIFICATION	REMARKS
	Or Eon To/Thora	Of Eon To/TiOT	Of Lon to thort	



Ref: BI-K/E-TEND/16/2019-20

Technical Specification of Glove Box

S.No.	Specification	Description
1	Glove box	 Internal dimensions (mm): 600 (W) x 450 (H) x 500 (D) (± 5%) Material of construction: 25 mm polycarbonate/acrylic transparent Environment: Nitrogen, Argon & Helium
2	Front view window/glass	Type: Scratch and chemical resistant glass or appropriate material, details should be provided
3	Shelves	Minimum 3 height adjustable, full length stainless steel shelves
4	Feed through (leak proof)	Electrical: Minimum 1, with minimum 1 outlet power strip
5	Filter	0.1-0.5 micron HEPA inlet/outlet filters (class H13 or better)
6	Main Anti Chamber	 Total: One Material of construction: 25 mm P.C/ acrylic transparent Door sealing mechanism: Spindle lock/ Gas piston type or appropriate
		Dimensions (mm): Diameter (300x300x300 mm)
7	Glove ports	 Total: Two Port material: Chemical resistant material details should be provided Diameter: 200 mm (± 10%), Sealing: O ring sealed
8	Gloves	Material: Chemical resistant Butyl or any appropriate, details should be provided Thickness: 0.4-0.6 mm (± 10%) Size: medium or large
9	Vacuum pump	Type1: Rotary vane mechanical pump for glove box pumping • Double stage • Base pressure without load 2x10 ⁻³ mbar • Flow volume: 15 m ³ /hr

		Type 2: Turbo molecular pump for sample transfer through vacuum jacket Inlet flange: DN 40 ISO-K Compression ratio for Nitrogen and Argon: > 1x10 ¹¹ Maximum backing vacuum pressure for nitrogen: > 20 mbar Normal rotation speed: 90,000 rpm or higher Gas throughput for Nitrogen: ≥ 1.3 mbar I/s Mounting Orientation: Any Ultimate vacuum with rotary vane baking pump: < 1x10 ¹⁷ mbar Vent or purge: automatic Cooling (35°C ambient) of turbo pump: Forced air cooling (optional) Bearing: dry permanent magnetic bearing at high vacuum side; oil lubricated ceramic bearing at other side Turbo controller: Suitable controller to power the turbo pump Vacuum interlock: Safety interlocks to stop rotation at the loss of backing vacuum Safety features: Stop rotation when the turbo is over heated Power: single phase AC 220 V 50 Hz Backing Pump: Rotary pump Peak pumping speed: ≥ 0.5 m³h⁻¹ Power: single phase AC 220 V 50 Hz Should be mounted on single frame Compound Vacuum Gauge: Pressure range: Atmosphere to 5x10⁻9 mbar Embedded gauges: Pirani and magnetron Accuracy: 30% < 100 mbar Gauge controller: inbuilt in the turbo controller Vacuum connector: NW25
		Vacuum connector: NW25 Power: single phase AC 220 V 50 Hz
10	Flow piping and fittings	Should be made of stainless steel/ silicon
11	Control system	Digital controlled of all basic functions & monitoring of box parameters with 6 inch (± 10%) monochrome/colour panel. Details of controls and functioning should be provided.
12	Warranty	Two years