

BOSE INSTITUTE KOLKATA

Tender No.	:	BI/Bio/SS/05/2013-14/R/3447/13
Tender date	:	28.10.2013
Type of Tender	:	Limited

Tender Title	:	Speed Vacuum Concentrator
Specification	:	<p>The Centrifugation system should be with timer, start stop function.</p> <p>Adjustable chamber temperature ambient +5 – 85 °C.</p> <p>Facility for digital read out of vacuum down to 1 mBar.</p> <p>RPM 0-2000 rpm.</p> <p>Automatic door release.</p> <p>Teflon membrane pump to 7 mBar or better, pumping capacity : 37 LPM or higher, condensate/drip catcher.</p> <p>Ice cube container unit to increase the performance and all connectors, hoses. Clips and Rotor 48 mL x 1.5-2.0 mL microfuge tubes.</p> <p>All accessories should be completed from one manufacturer.</p>
Quantity	:	01 No.

Last date & time for submission	:	15.11.2013 within 1.00 p.m.
Date & time for opening of bids	:	18.11.2013 at 2.00 p.m.
Submission of Tender (Address)	:	Office of the Department of Biochemistry, Bose Institute P-1/12, CIT Scheme VIIM, Kolkata 700054
Venue of bid opening	:	As above
For any query the interested bidders may contact (Dept./Section/Div./Unit)	:	Office of the Department of Biochemistry, Bose Institute 033 2569 3238

General Terms & Conditions ::		
Warranty	:	02 years
Payment terms	:	Payment will be made after complete delivery of the instrument in good condition and satisfactorily installation
Delivery schedule	:	Within 8 weeks from date of order & if any defect of the supplied item is found, it should be replaced immediately from your side.

Bid security (earnest money deposit), if applicable	:	N.A.
Submission of Performance Bank Guarantee (PBG), if applicable	:	N.A.
Any other information (if applicable)	:	
<i>Name of the instrument and submission of tender should be mentioned on the envelope positively.</i>		
Director, Bose Institute reserves the right to accept or reject any or all tenders either in part or in full. The reasons for rejecting the tender of a prospective bidder will be disclosed only when enquiries are made.		

Registrar